



Type examination certificate

No. CH-MI001-09027-01

Bauartprüfzertifikat

Applicant:
Auftraggeber:

G2 MISURATORI SRL
Via san Martino 38
14100 Asti
Italy

Requirements:
Anforderungen:

Swiss ordinance on measuring instruments (SR
941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and
Council of March 31, 2004 on measuring instruments
(MID), appendix B, class of measuring instruments
MI-001.

Conformity of superseded standard:
EN 14154 Edition 2007

Type of instrument:
Geräteart:

**Multi-jet impeller meter intended for the metering of
cold water**

Type designation:
Typenbezeichnung:

**AFM / AFM-P
BFM / BFM-P
PFM / PFM-P**

Accuracy class(es):
Genauigkeitsklasse(n):

2 (according to OIML R 49:2006)

Characteristics:
Kenndaten:

PN 16, DN 15 / 20 / 25 / 32 / 40 / 50, T30, T50, Class B, E1
Q₃ 1.6 / 2.5 / 4.0 / 6.3 / 10.0 / 16.0 / 25.0

Certificate valid until:
Zertifikat gültig bis:

July 22, 2019

CH-3003 Bern-Wabern, November 10, 2009

Notified body:
Benannte Stelle:

Certificate authority METAS-Cert
No. 1259

For the test:
Für die Prüfung:


Dr. Hugo Bissig


Jürg Ramseyer, Head of METAS-Cert